

Reference Window: 2.00 Minutes  
Non-Reference Window: 1.00 Minutes  
Correlation Window: 0.10 minutes  
Default Multiplier: 1.00  
Default Sample Concentration: 0.00

Compound Information

1) Methyltestosterone

(ISTD TR)

Ret. Time 20.92 min., Extract & Integrate from 20.42 to 21.42 min.

Signal Rel Resp. Pct. Unc.(rel) Integration  
Tgt 301.30 man27.e

Lvl ID	Conc (ng/mL)	Response
1	100.000	4680010
2	100.000	4212735
3	100.000	5428625

Qualifier Peak Analysis ON ISTD conc: 100.000 ng/mL  
Curve Fit: Linear

2) Epitestosterone

( )

Ret. Time 18.51 min., Extract & Integrate from 18.01 to 19.01 min.

Signal Rel Resp. Pct. Unc.(rel) Integration  
Tgt 432.40 man27.e

Lvl ID	Conc (ng/mL)	Response
1	5.000	181309
2	30.000	1100720
3	60.000	3350917

Qualifier Peak Analysis ON  
Curve Fit: Linear, forced through origin

3) testosterone

( )

Ret. Time 19.31 min., Extract & Integrate from 18.81 to 19.81 min.

Signal Rel Resp. Pct. Unc.(rel) Integration  
Tgt 432.40 man27.e

Lvl ID	Conc (ng/mL)	Response
1	30.000	1359912
2	180.000	7964015
3	360.000	21495301

Qualifier Peak Analysis ON